



Total Laboratory Services Ltd

Inspection Certificate

Dynamic calibration system, type AutoCal 2

Instrument No:

466698998

Sensor No:

734380

Inspection Date:

05-06-2025

Inspector:

P. Netherway

Measurement:	Nominal	Value	Result
Check 1:	10.0 kg \pm 0.1kg	10.0kg	Passed
Check 2:	20.0 kg \pm 0.1kg	20.0kg	Passed
Check 3:	30.0 kg \pm 0.1kg	30.0kg	Passed

Used measuring equipment (traceable to national standards):


Weight 20KG

W2404

Weight 10KG

W2405

Blandford Forum, Dorset



Certified by:

P. Netherway

Paul Netherway, Managing Director

Total Laboratory Services Ltd
Unit14 C, Sunrise Business Park
Blandford Forum, Dorset DT11 8ST