

CERTIFICATE of CALIBRATION

04 June 2025

Certificate No. YM 71876

Customer Total Laboratory Services Ltd
Description Digital Thermometer + Probe
Manufacturer TESTO
Model 720
Serial No. 03313682

Results:-

Customer Range	Applied Temperature °C	Indicator Reading °C
20°C to 50°C	20.00	20.1
	30.00	30.3
	40.00	40.3
	50.00	50.2

Note: 1 x Temperature Probe Satisfactory

New Battery Fitted

Uncertainty of Measurement +/- 0.6°C + 1 digit

We certify that the instrument detailed above has been calibrated at 20°C ±2°C.
The instrument was calibrated to known values through simulator.

All the equipment used in the calibration is certified and traceable to National Standards.

Traceability: 0961 UKAS 0601:281124/TCS3

Authorised Signatories:

A Lota ☐ K Burrow ☒ J Hatwell ☐

