

Certificate Number 48564

Page 1 of 2
08 July 2025

Total Laboratory Services Limited
Unit 14C
Sunrise Business Park
Blandford Forum
Dorset
DT11 8ST

CERTIFICATE OF CALIBRATION

We hereby certify that the 50g 20g 10g and 5g stainless steel listed weights in individual tube cases have been calibrated to fall within O.I.M.L. class F₁ tolerance.

Nominal Value	Measured Value g	Uncertainty +/- mg on UKAS Weights
50g (WT1939)	50.000 15	0.03
20g (WT1940)	20.000 09	0.025
10g (WT1941)	10.000 03	0.02
5g (WT1942)	5.000 09	0.015

The measured values reported in this certificate were determined by comparison weighing methods against our laboratory's reference standards with a hypothetical density of 8000 kg/m³ which in air of density 1.2kg/m³ would balance the nominal weight.

Traceability to National Standards is established by comparison to WEIGHTS standard class E₂ weight set number 988 certified on UKAS certificate number UP1228. Date of issue 21st August 2024 by Norfolk Calibration Services. UKAS calibration number 0260.

Recommended Recalibration - October 2026.

Why should recalibration be carried out? Recalibration of test equipment is a major requirement for quality management systems. All test weights vary with time due to wear and the collection of grime. The extreme of weight change varies with the environment the weights are used in, consequently periodic recalibration at regular intervals is required.

Signed



Weights Ltd

Weights Limited
27 Acacia Close, Cherrycourt Way
Leighton Buzzard, Bedfordshire
LU7 4QE

Tel: 01525 810919
sales@weightsltd.co.uk
www.weightsltd.co.uk

Certificate Number 48564

Page 2 of 2
08 July 2025

OIML +/- maximum permissible tolerance on calibration weights

Nominal Value	Class E ₁ mg	Class E ₂ mg	Class F ₁ mg	Class F ₂ mg	Class M ₁ mg	Class M ₂ mg
50Kg	25	80	250	800	2,500	8,000
20Kg	10	30	100	300	1,000	3,000
10Kg	5.0	16	50	160	500	1,600
5kg	2.5	8.0	25	80	250	800
2kg	1.0	3.0	10	30	100	300
1kg	0.5	1.6	5.0	16	50	160
500g	0.25	0.8	2.5	8.0	25	80
200g	0.10	0.3	1.0	3.0	10	30
100g	0.05	0.16	0.5	1.6	5.0	16
50g	0.03	0.10	0.3	1.0	3.0	10
20g	0.025	0.08	0.25	0.8	2.5	8.0
10g	0.020	0.06	0.20	0.6	2.0	6.0
5g	0.016	0.05	0.16	0.5	1.6	5.0
2g	0.012	0.04	0.12	0.4	1.2	4.0
1g	0.010	0.03	0.10	0.3	1.0	3.0
500mg	0.008	0.025	0.08	0.25	0.8	2.5
200mg	0.006	0.020	0.06	0.20	0.6	2.0
100mg	0.005	0.016	0.05	0.16	0.5	1.6
50mg	0.004	0.012	0.04	0.12	0.4	-
20mg	0.003	0.010	0.03	0.10	0.3	-
10mg	0.003	0.008	0.025	0.08	0.25	-
5mg	0.003	0.006	0.020	0.06	0.20	-
2mg	0.003	0.006	0.020	0.06	0.20	-
1mg	0.003	0.006	0.020	0.06	0.20	-

