

CERTIFICATE OF CALIBRATION

ISSUED BY **LEE CALIBRATION SERVICES LTD.**

DATE OF ISSUE **29 Aug 2025**

CERT NUMBER **26535**



0635

LCES

LEE CALIBRATION SERVICES LTD.

Units 3 & 4 Rose Farm Business Park, Leicester, LE8 5QW

Telephone: 0116 237 4446

E. Mail: lee-calibration@btconnect.com

PAGE	OF	PAGES
1	2	

APPROVED SIGNATORY

N. C. McDonald

G. Lee

E. McDonald

Issued to

TOTAL LABORATORY SERVICES LTD.

Dorset. DT11 8ST.

Order Number

2025169

Description : M3 A used set of metric steel gauge blocks

Serial Number : 160542/07430/08550

Cust. Asset No.

Report:-

Basis of calibration is to the accuracy requirements of BS EN ISO 3650:1999 Grade 2

This set of gauge blocks was measured at an ambient temperature of 20°C ± 1°C. The gauge blocks were calibrated for length and variation by measurement at 5 points, the centre point and at each corner, and the centre point deviations from nominal size and reported on page 2.

The measured values satisfied the tolerances specified in BS EN ISO 3650:1999 Grade 2. This certificate relates only to the items calibrated on this certificate.

The uncertainty of measurement is specified on page 2.

The statement of conformity to the specification is made in accordance with simple/shared risk decision rule

(REF 106:2012 Section 8.2) where the expanded uncertainty is =< than the specification/tolerance

Certified by :- **E McDonald**

Comments:-

Traceability :-

Master Set no. :- **12095/157C**

Certificate No. :-

14239

Master Set Material :- **steel**

Master Expansion Coefficient : **11.5 ppm / °C**

Comparator Used :- **VC5 529**

Software Version :-

Slipsure 3.2.0.7

A blue ink signature scribble.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of unite and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

LFE CALIBRATION SERVICES LTD.

UKAS ACCREDITED CALIBRATION LABORATORY NO. 0635

Continuation

CERT NUMBER

26535

PAGE OF PAGES

2 of 2

Date of Calibration :- 29 Aug 2025

Serial Number : 160542/07430/08550

Cust. Asset No.

The centre measured deviation from nominal in units of 0.00001 mm

Nominal (mm)	Deviation
3.00	+8
5.00	+2
10.00	-14

Uncertainty of Measurement (\pm) Units 0.00001 mm

Gauges above 0 mm up to 10 mm 8

The measured values were within the tolerance specified in BS EN ISO 3650:1999 Grade 2

END OF REPORT

