



CERTIFICATE of CALIBRATION

03 October 2025

Certificate No. YM 73815

Customer Total Laboratory Services Ltd
 Description Digital Thermometer + Probe
 Manufacturer Greisinger
 Model GMH3710
 Serial No. 32408697

Results:-

Customer Range	Applied Temperature °C	Indicator Reading °C
	20.00	19.95
	30.00	29.96
	40.00	39.98
20°C to 50°C	50.00	49.99

Note: 1 x Temperature Probe Satisfactory

Uncertainty of Measurement +/- 0.6°C + 1 digit

We certify that the instrument detailed above has been calibrated at 20°C ±2°C.
 The instrument was calibrated to known values through simulator.

All the equipment used in the calibration is certified and traceable to National Standards.

Traceability: 0961 UKAS 0601:281124/TCS3

Authorised Signatories:

A Lota K Burrow J Hatwell