

# Inspection Certificate


## Distance Plates

Serial Number: V95005

Inspection date: 23-10-2025

Inspector: C. Juett

Thickness Measurement	Nominal	Actual		
		Average	Xmin	Xmax
Distance plate 0.6mm	0.6 ±0.05	0.60	0.58	0.61
Distance plate 1.0mm	1.0 ±0.05	0.99	0.95	1.03
Distance plate 1.5mm	1.5 ±0.05	1.50	1.47	1.52



Material: Stainless Steel

Measuring equipment\*: Caliper 019160

Dorset, 23.10.2025

i.A. *P. Netherway*  
Paul Netherway  
- Managing Director -

Our testing and measuring equipment are subject to measuring equipment monitoring and is audited annually in accordance with ISO 9001 guidelines

\*Traceable to national standards