



Total Laboratory Services Ltd

Certificate No: TLSCJUV006

## Certificate of Operational Qualification

**Instrument:** Shimadzu 1800 UV-Visible Spectrophotometer

**Serial No:** A11635271379

The above named instrument has been successfully qualified against the following tests and is valid for use until the period specified below.

Test	Requirement
Photometric Accuracy & Linearity (Potassium Dichromate)	USP/EP/JP ASTM & NIST
Wavelength Accuracy & Repeatability (Holmium Oxide)	USP/EP/JP ASTM
Stray Light Test (Potassium Chloride)	USP/EP/JP ASTM
Resolution (Toluene in Hexane)	USP/EP
Baseline Flatness	Manufacturer

**Result:** Passed

**Qualification date:** 20th October 2025


**Valid until:** 20th October 2026

**Tested by:**

Chris Juett, Service Engineer

**Approved by:**

Paul Netherway, Managing Director



Total Laboratory Services Ltd  
Unit 14 C, Sunrise Business Park  
Blandford Forum, Dorset DT11 8ST