

CERTIFICATE OF CALIBRATION

Issued by: **Compli Group Ltd.**

BS EN ISO9001:2015 LRQA Registered Company No. LRQ 00005310

Unit 1, Cabot Business Village, Holyrood Close, Poole, Dorset, BH17 7BA

www.compli.uk

Telephone: 01202 656165

Email: sales@compli.uk

Date of Issue: 04 December 2025

Certificate Number: 439211

Customer:

Total Laboratory Services (TLS)

Unit 14C

Sunrise Business Park

Blandford Forum

Dorset

DT11 8ST

Page 1 of 1 Page

Approved Signatory

A. Garnett



Equipment: Inclinometer Solatronic,

Serial No: 96021

Conditions of Calibration: <70 %rh & 20 °C ±2 °C

Date of Calibration: 04 December 2025 **Procedure No.** Pro 0050

Acceptance Criteria: Manufacturer's Specification (0 °/90 °=0.05 °, 1 ° to 89 °=0.2 °)

Method of Calibration: Calibrated using angle blocks and a surface table

Results: As found unless specified

Angle Applied	Incline Inclinometer Reading	Decline Inclinometer Reading	Tolerance ±	Result
0.0	0.0	0.0	0.05	Pass
1.0	1.0	1.0	0.20	Pass
3.0	3.0	2.9	0.20	Pass
9.0	8.9	8.9	0.20	Pass
27.0	26.9	26.9	0.20	Pass
41.0	40.9	40.9	0.20	Pass
90.0	90.0	-	0.05	Pass

Measurement Uncertainty: ±0.1 °

Measured results are traceable to national standards via equipment

PICL205, Certificate: 394843 (Cal due: 07/11/2026); PICL271, Certificate: 160499 (Cal due: 21/06/2027)

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%