

# CERTIFICATE OF CALIBRATION

Issued by: Compli Group Ltd.

BS EN ISO9001:2015 LRQA Registered Company No. LRQ 00005310

Unit 1, Cabot Business Village, Holyrood Close, Poole, Dorset, BH17 7BA

www.compli.uk

Telephone: 01202 656165

Email: sales@compli.uk

Date of Issue: 03 December 2025

Certificate Number: 439212

Customer:  
Total Laboratory Services (TLS)  
Unit 14C  
Sunrise Business Park  
Blandford Forum  
Dorset  
DT11 8ST

Page 1 of 1 Page  
Approved Signatory  
M. Hill



**Equipment:** Feeler Gauge Set, Metric,

**Serial No:** 2693856/324

**Conditions of Calibration:** <70 %rh & 20 °C ±2 °C

**Date of Calibration:** 03 December 2025 **Procedure No.** Pro 0065

**Acceptance Criteria:** BS 957:2008

**Method of Calibration:** Calibrated using a horizontal metroscope.

**Results:** As found unless specified

Feeler Indication mm	Measured Deviation 1=0.001 mm	Tolerance ± 1=0.001 mm	Result
0.05	-1.4	5.0	Pass
0.10	+1.2	5.0	Pass
0.15	+1.8	5.0	Pass
0.20	+1.6	5.0	Pass
0.25	+5.0	5.0	Pass
0.30	+0.4	8.0	Pass
0.40	+0.2	8.0	Pass
0.60	+2.4	8.0	Pass
0.80	+10.0	10.0	Pass
0.90	+0.9	10.0	Pass
1.00	+0.6	10.0	Pass

Measurement Uncertainty: +/- 0.002 mm

Measured results are traceable to national standards via equipment  
PICL077, Certificate: 399518 ( Cal due: 05/05/2026 )

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%