

CERTIFICATE OF CALIBRATION

Issued by: **Compli Group Ltd.**

BS EN ISO9001:2015 LRQA Registered Company No. LRQ 00005310

Unit 1, Cabot Business Village, Holyrood Close, Poole, Dorset, BH17 7BA

www.compli.uk

Telephone: 01202 656165
Email: sales@compil.uk

Date of Issue: 26 February 2026

Certificate Number: 445659

Customer:

Total Laboratory Services (TLS)
Unit 14C

Sunrise Business Park
Blandford Forum
Dorset
DT11 8ST

Page 1 of 1 Page
Approved Signatory
A. Garnett



Equipment: Inclinometer Messmittel,

Serial No: S93034

Conditions of Calibration: <70 %rh & 20 °C ±2 °C

Date of Calibration: 26 February 2026 **Procedure No.** Pro 0050

Acceptance Criteria: None

Method of Calibration: Calibrated using angle blocks on a surface table

Results: As found unless specified

Applied Angle	Inclinometer	Inclinometer
	Reading Incline	Reading Decline
0.0 °	0.0 °	0.0 °
1.0 °	1.0 °	1.0 °
3.0 °	3.0 °	3.0 °
9.0 °	9.0 °	9.0 °
27.0 °	26.9 °	27.0 °
41.0 °	41.0 °	41.0 °
90.0 °	90.0 °	-

Measurement Uncertainty: ±0.1 °

Measured results are traceable to national standards via equipment

PICL205, Certificate: 394843 (Cal due: 07/11/2026); PICL271, Certificate: 160499 (Cal due: 21/06/2027)

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%