


ARK



****Calibration Certificate****

Do Not Destroy


Calibration Certificate Attached: 1963249
OD ref: 1262590841

150 mm Digital Caliper

first

IMPORTANT INFORMATION

Simply detach the label in the top right hand corner of the new front sheet and apply to your instrument as required.



For Re-Calibration of your unit please email:

calibration.uk@rs-components.com

or call us on 01536 405545 to arrange free collection. Please quote serial number when returning.

RS Calibration

CERTIFICATE OF CALIBRATION

Issued by: RS Components Ltd

Date Issued: 25 Feb 2026

Certificate No. 1963249



Calibration and Repair Service

DPN 175, Lammas Rd,
Weldon Industrial Est
Corby, Northants, NN17 9RS

Tel: 01536 405545

Fax: 01536 401590

Client

TOTAL LABORATORY SERVICES LTD
BLANDFORD FORUM
DORSET
DT11 8ST

Instrument

150 mm Digital Caliper

Serial No.

019160

Client Reference

N/A

Procedure ID.

D02_1150_# Rev. P4(Code: SCQAR532DIG)

Date of Calibration

25 Feb 2026

Remarks

This certificate reports recorded values for the instrument 'As Received'.

Uncertainties

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

For certificate statements of conformity see Appendix SCQAR 533
The following calibration results relate only to the items defined above.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.



0310



Page 1 of 3 Pages

Ashida Kunnath

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0310



Calibration and Repair Service

Certificate No.

1963249

Page 2 of 3 Pages

Environment

The ambient temperature and relative humidity throughout the calibration were (20 ± 2) °C and (40 ± 35) %rh respectively.

Calibration

The instrument was held in a temperature controlled environment for a period of not less than 4 hours prior to calibration which was performed using gauge blocks, length bars and in accordance with RS Calibration procedure MLCP01. The results are compared with the tolerances specified by SCQAR532_RS Standard Limits For Calipers.

		Measurement Uncertainty
Flatness of external faces:	0.001 mm	+/- 0.005 mm
Parallelism of measuring jaws:		
Internal	0.021 mm	+/- 0.005 mm
External	<u>At 1 mm</u> 0.001 mm <u>At 140 mm</u> 0.001 mm	+/- 0.005 mm

No limits are applied to the above values.

Calibration of Gauge

	Tested Size (mm)	Limits (mm)	Gauge Reading (mm)	Deviation (mm)	Measurement Uncertainty (mm)
<u>External</u>	0.00	+/- 0.02	0.00	0.00	+/- 0.012
	31.20	+/- 0.02	31.19	-0.01	+/- 0.012 #
	61.40	+/- 0.02	61.40	0.00	+/- 0.012
	92.60	+/- 0.02	92.59	-0.01	+/- 0.013 #
	123.80	+/- 0.03	123.79	-0.01	+/- 0.014
	150.00	+/- 0.03	149.99	-0.01	+/- 0.015
<u>Internal</u>	50.00	+/- 0.04	50.02	0.02	+/- 0.012
<u>Step</u>	50.00	+/- 0.04	49.99	-0.01	+/- 0.012
<u>Depth Rod</u>	50.00	+/- 0.04	50.00	0.00	+/- 0.012

Calibrated by: ARK

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0310



Calibration and Repair Service

Certificate No.

1963249

Page 3 of 3 Pages

Reported values not annotated.

The instrument passed the stated specification, due allowance having been made for the uncertainty of measurement which carries no implication regarding the long term stability of the instrument.

Reported values annotated with a #

The measured result is a conditional pass to the specification limit by a margin less than the measurement uncertainty; it is therefore not possible to state compliance based on the stated level of confidence.

END OF CALIBRATION

Appendix SCQAR533 Certificate Statements of conformity

RS Components is standardising how it reports conformity across all disciplines in line with requirements within **ISO/IEC: 17025:2017**.

Where the laboratory reports a statement of conformity to a specification, guidance has been drawn on reporting structure and decision rules from ILAC document series **ILAC-G8:09/2019**.

Unless otherwise instructed by you the Customer, acceptance limits applied are derived from the manufacturers specification or applicable standard (e.g. DIN, EEC, BS etc.) or where applicable: SCQAR532_RS Standard Limits for Calipers, available on request.

The statements found on this certificate produced by RS Components Laboratory are as follow:

1) Reported values with **No Annotation**:

The instrument **passed** the stated specification, even with allowance having been made for the uncertainty of measurement, which carries no implication regarding the long-term stability of the instrument.

2) Reported values annotated with **"#"**

The measured result is a **conditional pass** to the limit but by a margin less than the measurement uncertainty, it is therefore not possible to state compliance based on the stated level of confidence.

3) Reported values annotated with **"###"**

The measured result is a **conditional fail** to the limit but by a margin less than the measurement uncertainty, it is therefore not possible to state compliance based on the stated level of confidence.

4) Reported values annotated with **"###"**

The measured result **failed** the stated specification, even with allowance having been made for the measurement uncertainty.

