

**** Calibration Certificate **** **Do Not Destroy**

Calibration Certificate Attached: 1974961

OD ref: 1265923414

Stainless Steel Rule, Class 2, 150mm

first



IMPORTANT INFORMATION

Simply detach the label in the top right hand corner of the new front sheet and apply to your instrument as required.



For Re-Calibration of your unit please email:

calibration.uk@rs-components.com

or call us on 01536 405545 to arrange free collection. Please quote serial number when returning.

RS Calibration

CERTIFICATE OF CALIBRATION

Issued by: RS Components Ltd

Date Issued: 08 May 2026

Certificate No. 1974961



0310

RS Calibration

Calibration and Repair Service

DPN 175, Lammas Rd,
Weldon Industrial Est
Corby, Northants, NN17 9RS

Tel: 01536 405545
Fax: 01536 401590

Page 1 of 4 Pages

Paul Frost

Tel: 01536 405545
Fax: 01536 401590

Client

TOTAL LABORATORY SERVICES LTD
BLANDFORD FORUM
DORSET
DT11 8ST

Instrument

Stainless Steel Rule, Class 2, 150mm

Serial No.

87606

Client Reference

N/A

Procedure ID.

D05_1200_# Rev. P 8

Date of Calibration

08 May 2026

Remarks

This certificate reports recorded values for the instrument 'As Received'.

Item	Value	Units	Uncertainty	Pass/Fail
1	150.00	mm	±0.02	Pass
2	150.00	mm	±0.02	Pass
3	150.00	mm	±0.02	Pass
4	150.00	mm	±0.02	Pass
5	150.00	mm	±0.02	Pass
6	150.00	mm	±0.02	Pass
7	150.00	mm	±0.02	Pass
8	150.00	mm	±0.02	Pass
9	150.00	mm	±0.02	Pass
10	150.00	mm	±0.02	Pass
11	150.00	mm	±0.02	Pass
12	150.00	mm	±0.02	Pass
13	150.00	mm	±0.02	Pass
14	150.00	mm	±0.02	Pass
15	150.00	mm	±0.02	Pass
16	150.00	mm	±0.02	Pass
17	150.00	mm	±0.02	Pass
18	150.00	mm	±0.02	Pass
19	150.00	mm	±0.02	Pass
20	150.00	mm	±0.02	Pass

Uncertainties

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

The uncertainty evaluation has been carried out in accordance with UKAS requirements.

For certificate statements of conformity see Appendix SCQAR 533

The following calibration results relate only to the items defined above.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0310

Certificate No.
1974961



Calibration and Repair Service

Page 2 of 4 Pages

Environment

Prior to calibration the rule was held within a temperature controlled environment for a period of not less than 4 hours.

The ambient temperature and relative humidity throughout the calibration were (20 ± 2) °C and (40 ± 35) %rh respectively.

Method

The scale identified below was calibrated by measuring from the edge of the rule to the first position. This first position was then used as a datum from which all other positions on that scale are referenced. Measurements were made using a horizontal length measuring machine and the results recorded in the tables below.

The calibration was performed in accordance with 73-362 / EEC Class 2.

Side One

Top Scale

Major Position	Nominal Length	Measured Length	Measured Deviation	Major Position Limits	Measurement Uncertainties
0 - 10	10	9.970	-0.030	± 0.500	± 0.009
10 - 30	20	19.978	-0.022	± 0.500	± 0.009
10 - 31	21	20.994	-0.006	± 0.500	± 0.009
10 - 59	49	48.981	-0.019	± 0.500	± 0.009
10 - 60	50	49.986	-0.014	± 0.500	± 0.009
10 - 90	80	79.965	-0.035	± 0.500	± 0.009
10 - 91	81	80.978	-0.022	± 0.500	± 0.009
10 - 119	109	108.953	-0.047	± 0.500	± 0.010
10 - 120	110	109.950	-0.050	± 0.500	± 0.010
10 - 150	140	139.938	-0.062	± 0.500	± 0.010
Adjacent Position	Nominal Length	Measured Length	Measured Deviation	Adjacent Position Limits	Measurement Uncertainties
mm	mm	mm	mm	mm	mm
30 - 31	1	1.016	0.016	± 0.200	± 0.009
59 - 60	1	1.005	0.005	± 0.200	± 0.009
90 - 91	1	1.013	0.013	± 0.200	± 0.009
119 - 120	1	0.997	-0.003	± 0.200	± 0.009

Side One

Maximum deviation found between any two major positions in the above table from 10mm to the maximum length.

Major position limit

Measurement Uncertainty

-0.062 mm
± 0.500 mm
± 0.010 mm

Maximum deviation found between any adjacent positions.

Adjacent position limit

Measurement Uncertainty

0.016 mm
± 0.200 mm
± 0.009 mm

Appendix SCQAR533 Certificate Statements of conformity

RS Components is standardising how it reports conformity across all disciplines in line with requirements within ISO/IEC: 17025:2017.

Where the laboratory reports a statement of conformity to a specification, guidance has been drawn on reporting structure and decision rules from ILAC document series ILAC-G8:09/2019.

Unless otherwise instructed by you the Customer, acceptance limits applied are derived from the manufacturers specification or applicable standard (e.g. DIN, EEC, BS etc.) or where applicable: SCQAR532_RS Standard Limits for Calipers, available on request.

The statements found on this certificate produced by RS Components Laboratory are as follow:

1) Reported values with No Annotation:

The instrument **passed** the stated specification, even with allowance having been made for the uncertainty of measurement, which carries no implication regarding the long-term stability of the instrument.

2) Reported values annotated with “#”

The measured result is a **conditional pass** to the limit but by a margin less than the measurement uncertainty, it is therefore not possible to state compliance based on the stated level of confidence.

3) Reported values annotated with “###”

The measured result is a **conditional fail** to the limit but by a margin less than the measurement uncertainty, it is therefore not possible to state compliance based on the stated level of confidence.

4) Reported values annotated with “####”

The measured result **failed** the stated specification, even with allowance having been made for the measurement uncertainty.



