



Right Service - Right Solution

Certificate No: CJ131026161456UV11

Certificate of Operational Qualification

Instrument: Shimadzu 1800 UV-Visible Spectrophotometer

Serial No: A11635271379

The above named instrument has been successfully qualified against the following tests and is valid for use until the period specified below.

Test	Requirement
Photometric Accuracy & Linearity (Potassium Dichromate)	USP/EP/JP ASTM & NIST
Wavelength Accuracy & Repeatability (Holmium Oxide)	USP/EP/JP ASTM
Stray Light Test (Potassium Chloride)	USP/EP/JP ASTM
Resolution (Toluene in Hexane)	USP/EP
Baseline Flatness	Manufacturer

Result: Passed

Qualification date: 13th October 2026

Valid until: 13th October 2027

Tested by:

Chris Juett, Service Engineer

Approved by:

Paul Netherway, Managing Director

